



HISCL™ FT4 Calibrator

REF 054-2381-2

Identification of the IVD reagent

HISCL™ FT4 Calibrator

Intended use

For In Vitro Diagnostic Use

For calibration of HISCL FT4 Assay Kit

Components

- 1) HISCL FT4 C0
- 2) HISCL FT4 C1
- 3) HISCL FT4 C2
- 4) HISCL FT4 C3
- 5) HISCL FT4 C4
- 6) HISCL FT4 C5

Warnings and precautions

- 1) All calibrators contain human serum. Handle with care as potentially infectious, though the calibrator material has been checked as negative for HBs antigen, HCV antibody, HIV-1 antibody, and HIV-2 antibody.
- 2) Each calibrator contains sodium azide. Since sodium azide reacts with lead tubing and copper tubing to generate metal azides which can explode, use plenty of water when disposing it. In case of contact with the eyes, mouth, or hands, carry out emergency treatment such as washing with plenty of water. If necessary, consult a physician.

Examination procedure

Dispense about 200µL (5 drops) of each calibrator.

Carry out procedures according to the package insert for the HISCL FT4 Assay Kit.

Storage and shelf life after first opening

Store at 2-8°C. The shelf life is 90 days after opening.

Do not freeze.

Manufacturer



Sysmex Corporation

1-5-1 Wakinohama-Kaigandori, Chuo-ku, Kobe 651-0073, Japan

Authorized representative / Distributors

Asia-Pacific: **Sysmex Asia Pacific Pte Ltd.**

9 Tampines Grande #06-18, Singapore 528735

Date of issue or revision

09/2015

Assigned Value

Assigned values are indicated on the reverse side of this document.

Set these values according to the instruction manual for the instrument used.

Symbol

Refer to the HISCL operator's manual for the details of the symbols written on the outer box.

Sysmex Corporation

Printed in Japan
54505097K