

REF AT-609-958

HISCL™ M2BPGi™ Calibrator

Identification of the IVD reagent**HISCL™ M2BPGi™ Calibrator****Intended use**

For In Vitro Diagnostic Use

For calibration of HISCL M2BPGi Assay Kit.

Components

1. HISCL M2BPGi NC (M2BPGi Negative Calibrator)
2. HISCL M2BPGi PC (M2BPGi Positive Calibrator)

Warnings and precautions

Each calibrator contain sodium azide. Since sodium azide reacts with lead tubing and copper tubing to generate metal azides which can explode, use a large quantity of water when disposing it. In case of contact with the eyes, mouth, or hands, perform emergency treatment such as washing with a large quantity of water. If necessary, consult a physician.

Examination procedure

Dispense about 200 µL (5 drops) of each calibrator.

Carry out procedures according to the package insert for the HISCL M2BPGi Assay Kit.

Storage and shelf life after first opening

Store at 2-8°C. The shelf life is 90 days after opening.

Do not freeze.

Manufacturer**Sysmex Corporation**

1-5-1 Wakinohama-Kaigandori, Chuo-ku, Kobe 651-0073, Japan

Authorized representative / DistributorsAsia-Pacific: **Sysmex Asia Pacific Pte Ltd.**

9 Tampines Grande #06-18, Singapore 528735

Traceability of values assigned to calibrators and trueness-control materials

HISCL M2BPGi PC has been adjusted by in-house standard materials.

Date of issue or revision

01/2016

Symbol

Refer to the HISCL Instructions for Use for the details of the symbols written on the outer box.

Printed in Japan