

# Dade<sup>®</sup> Actin<sup>®</sup> Activated Cephaloplastin Reagent

**ACTIN**

I Revision bar indicates update to previous version.

**C€0197**

## Intended Use

**ACTIN** is an in vitro diagnostic reagent for the quantitative determination of activated partial thromboplastin time (APTT) as an aid to diagnosis, screening for hemostasis disorders and monitoring of unfractionated heparin in human sodium citrated plasma by means of automated, semi-automated and/or manual coagulometric methods.

For APTT testing no international reference preparation or method is available.

## Summary and Explanation

**ACTIN** is a liquid reagent based on rabbit brain cephalin with ellagic acid for plasma activation.

The APTT, a global screening procedure<sup>1,2</sup> used primarily to evaluate coagulation abnormalities in the intrinsic pathway, will also detect severe functional deficiencies in factors: FII, FV, FX, or fibrinogen. The APTT has also been widely advocated<sup>3-6</sup> as a means to monitor the effectiveness of unfractionated heparin therapy where the clotting time is prolonged in proportion to the level of heparin. In patients receiving oral anticoagulants, the circulating levels of factors: FII, FV, FIX, and FX are depressed therefore the APTT can be expected to be prolonged. The presence of non-specific inhibitors, such as the lupus-like anticoagulant<sup>1,7</sup>, may prolong the APTT but this effect is variable and generally recognized as being related more to the nature of the APTT reagent employed. In summary, the APTT is a clinically important screening test with wide applicability for the diagnosis of coagulant disorders and therapeutic monitoring of both, hemorrhagic and thrombotic disease<sup>8-10</sup> and specifically used

- In pre-surgery bleeding risk assessment
- In screening for bleeding disorders, e.g. in case of suspected intrinsic factor deficiency or inhibitors to intrinsic coagulation factors
- As an aid to diagnosis of lupus anticoagulants (inhibitors) in patients with thrombophilia
- For monitoring of therapy in patients receiving unfractionated heparin

Furthermore, **ACTIN** can be used in combination with the respective factor FVIII, FIX, FXI or FXII deficient plasma for the quantification of the coagulation factors: FVIII, FIX, FXI and FXII.

## Principles of the Procedure

Factors of the intrinsic coagulation system are activated by incubating the plasma with the optimal amount of phospholipids and a surface activator. The addition of calcium ions triggers the coagulation process, and the clotting time is then measured.

## Reagents

**Note:** **ACTIN** can be used on automated coagulation analyzers. Sysmex provides Reference Guides (Application Sheets) for several coagulation analyzers. The Reference Guides (Application Sheets) contain analyzer/assay specific handling and performance information which may differ from that

provided in these Instructions for Use. In this case, the information contained in the Reference Guides (Application Sheets) supersedes the information in these Instructions for Use. Please also consult the instruction manual of the instrument manufacturer!

| Reagent                                                       | Description                                                                                                                                                                                                           | Storage                                                                                                  | Stability                                    |
|---------------------------------------------------------------|-----------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------|----------------------------------------------------------------------------------------------------------|----------------------------------------------|
| Dade® Actin® Activated Cephaloplastin Reagent<br><b>ACTIN</b> | Ready to use liquid containing: <ul style="list-style-type: none"><li>rabbit cephaline<sup>a</sup> in <math>1.0 \times 10^{-4}</math> M ellagic acid</li><li>Preservative</li><li>Stabilizer</li><li>Buffer</li></ul> | 2–8 °C<br>May be used up to the expiry date indicated on the label if stored unopened.<br>Do not freeze! | 2–15 °C:<br>once opened, 7 days <sup>b</sup> |

<sup>a</sup> Standard activity values for rabbit cephalin are not available.

<sup>b</sup> closed original vial

If the reagent is left to stand, a green deposit may form consisting of ellagic acid and lipids. Before use, mix by inverting. Avoid contamination with plasma.

**Signs of expiry:** Deviations from the normal laboratory value in the determination of normal plasma or controls.

#### **On-board stability**

Information regarding on-board stability is specified in the Reference Guides (Application Sheets) for the different coagulation analyzers.

## **Warnings and Precautions**

For *in-vitro* diagnostic use only.

For laboratory professional use.

According to EU regulation 2017/746, any serious incident that has occurred in relation to the device shall be reported to the manufacturer and the competent authority of the EU Member State through your local distribution representative in which the user and/or patient is established.

Safety data sheets (MSDS/SDS) available upon request.

#### **Caution**

This device contains material of animal origin and should be handled as a potential carrier and transmitter of disease.

Dispose of hazardous or biologically contaminated materials according to the practices of your institution. Discard all materials in a safe and acceptable manner and in compliance with all government requirements.

Summary of Safety and Performance (SSP) is available in the European database on medical devices (see Eudamed public website: <https://ec.europa.eu/tools/eudamed>). In case Eudamed is not available, SSP can be delivered by the manufacturer on request.

## **Preparing Reagents**

**ACTIN** must be mixed gently by inversion (5 to 8 times) before use.

## **Specimen Collection and Handling**

### **Collecting the Specimen**

Mix nine parts of freshly collected patient blood with one part of 0.11 mol/L or 0.13 mol/L (3.2 % or 3.8 %) sodium citrate<sup>24,25</sup>. It is recommended that blood specimens for plasma-based coagulation testing should be collected by venipuncture using a blood collection system that collects the specimen directly into a glass or plastic evacuated tube containing the appropriate additive. Evacuated tubes containing the desired anticoagulant are commercially available and may be used with caution in blood coagulation studies.

For special studies, syringe technique may be preferred.

Centrifuge the blood specimen at  $1500 \times g$  for no less than 15 minutes at room temperature as soon as possible after collection.

## Storing the Specimen

Store in an unopened tube at room temperature.

If immediate testing is to be done, the plasma may remain on the packed cells. Otherwise plasma should be separated from the cells. To separate the plasma, use a plastic transfer pipette, remove plasma to a plastic tube.

Do not store on ice.

Non-heparinized plasma should be tested within 4 hours of blood collection.

Plasma containing unfractionated heparin should be centrifuged within one hour of blood collection, stored at room temperature and tested within 4 hours<sup>26</sup>.

Platelet-poor plasma may be frozen at  $\leq -20$  °C for up to 2 weeks in a non frost-free freezer.

Frozen plasma should be rapidly thawed at 37 °C, gently mixed and tested immediately. Samples should not stand at 37 °C for >5 minutes<sup>27</sup>.

Please refer to CLSI document H21-A58<sup>11</sup> for detailed information on sample preparation and storage.

## Procedure

### Materials Provided

| REF     | Contents                                               |      |       |
|---------|--------------------------------------------------------|------|-------|
| B4218-1 | Dade® Actin® Activated Cephaloplastin Reagent<br>ACTIN | 10 × | 2 mL  |
| B4218-2 | Dade® Actin® Activated Cephaloplastin Reagent<br>ACTIN | 10 × | 10 mL |

### Materials Required but not Provided

| Item         | Description                                                                                                                            |
|--------------|----------------------------------------------------------------------------------------------------------------------------------------|
| REF ORHO37   | CaCl <sub>2</sub> SOLUTION, Calcium Chloride Solution, (0.025 mol/L)                                                                   |
| REF ORKE41   | CONTROL N, Control Plasma N, or                                                                                                        |
| REF 291070   | Dade® Ci-Trol® 1, or                                                                                                                   |
| REF B4244-10 | Ci-Trol CONTROL 1, Dade® Ci-Trol® Coagulation Control Level 1, as control for the normal range                                         |
| REF OUPZ17   | CONTROL P, Control Plasma P, or                                                                                                        |
| REF 291071   | Dade® Ci-Trol® 2, or                                                                                                                   |
| REF B4244-20 | Ci-Trol CONTROL 2, Dade® Ci-Trol® Coagulation Control Level 2, as control for the pathological/therapeutical range                     |
| REF 291072   | Dade® Ci-Trol® 3, or                                                                                                                   |
| REF B4244-30 | Ci-Trol CONTROL 3, Dade® Ci-Trol® Coagulation Control Level 3, as control for the pathological/therapeutical range                     |
| REF B4224-50 | Ci-Trol HEPARIN CONTROL 1, Dade® Ci-Trol® Heparin Control, Low                                                                         |
| REF B4224-60 | Ci-Trol HEPARIN CONTROL 2, Dade® Ci-Trol® Heparin Control, High                                                                        |
| –            | For blood collection, use sodium citrate (0.11 mol/L or 0.13 mol/L/3.2 % or 3.8 %), or<br>Standard commercial blood collection systems |
| –            | Distilled or deionized water without preservatives                                                                                     |
| –            | Plastic test tubes                                                                                                                     |

| Item                                          | Description                                                                                                                                                                                                                                                                                                                                               |
|-----------------------------------------------|-----------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------|
| –                                             | Pipettes for precise measurement of 0.1 mL                                                                                                                                                                                                                                                                                                                |
| Coagulation analyzers <sup>c</sup> , such as: | <ul style="list-style-type: none"> <li>Automated Blood Coagulation Analyzer CA-600 series (CA-600 series)</li> <li>AUTOMATED BLOOD COAGULATION ANALYZER CS-2500 (CS-2500 System)</li> <li>AUTOMATED BLOOD COAGULATION ANALYZER CS-5100 (CS-5100 System)</li> <li>Automated Blood Coagulation Analyzer CN-3000/CN-6000 (CN-3000/CN-6000 System)</li> </ul> |

<sup>c</sup> Availability of analyzers may vary by country.

Please note that the applications on other analyzers can be validated by the instrument manufacturer in accordance with the requirements of the REGULATION (EU) 2017/746 under their responsibility as long as the intended purpose and performance are not modified.

## Manual Testing

| Pre-warm CaCl <sub>2</sub> <b>SOLUTION</b> at 37 °C                                                                                                |             |                |
|----------------------------------------------------------------------------------------------------------------------------------------------------|-------------|----------------|
| Pre-warm 0.1 mL <b>ACTIN</b> for 1 minute at 37 °C. (Mix before use)                                                                               |             |                |
| Pipet into coagulation tubes as follows:                                                                                                           |             |                |
|                                                                                                                                                    | Test Sample | Control Plasma |
| <b>ACTIN</b> (pre-warmed)                                                                                                                          | 0.1 mL      | 0.1 mL         |
| Plasma                                                                                                                                             | 0.1 mL      | –              |
| Control Plasma                                                                                                                                     | –           | 0.1 mL         |
| Mix well. Incubate at 37 °C for 3 minutes.                                                                                                         |             |                |
| CaCl <sub>2</sub> <b>SOLUTION</b> (pre-warmed)                                                                                                     | 0.1 mL      | 0.1 mL         |
| Simultaneously with addition of CaCl <sub>2</sub> <b>SOLUTION</b> start stopwatch, mix well. After 20 seconds start to observe for clot formation. |             |                |

**Note:** Incubation times exceeding 5 minutes may cause loss of FV and FVIII and are not recommended.

Each laboratory should determine the optimal heating-activation time for its particular assay system.

## Monitoring of Unfractionated Heparin Therapy with APTT

When using the APTT for this purpose, the factors influencing the test should be kept in mind. General considerations are listed below:

- Time of collection is important since the *in-vivo* half-life of unfractionated heparin is approximately 1.5 hours<sup>5</sup>. When it is administered, it has an immediate anticoagulant effect but the degree of this effect decreases rapidly with time. This is especially apparent with intermittent single intravenous injections.
- The anticoagulant used for sample collection can alter test results.
- Platelet factor 4, a heparin neutralizing factor in platelet alpha-granules, can be released by platelet aggregation or damage. To prevent this occurrence *in-vitro*, the specimen should be collected with a minimum of trauma. Cold temperatures are known to induce platelet aggregation and release platelet factor 4; therefore, centrifugation at room temperature is recommended for heparin studies.
- Using APTT to monitor unfractionated heparin therapy is time-dependent. Delay in testing samples will result in prolonged APTT determinations. Therefore, it is imperative that the testing on all samples be performed as soon as possible.
- Increased contact activation times may result in prolonged APTT in plasma containing heparin. It is imperative that the optimal heating-activation time of the **ACTIN**-plasma mixture be rigidly standardized<sup>12</sup>.
- Different test systems (i.e. manual, photo-optical, etc.) will exhibit variable heparin sensitivity. Interchanging of test systems should be avoided.

- G. Baseline data on the APTT of each patient before the start of therapy should be established where feasible to determine the respective patient APTT as it relates to the normal range established for the test in that laboratory.
- H. Studies have shown variability in original estimates of the quality of unfractionated heparin from different sources and different manufacturers. *In-vivo* reactivity varies with the type of heparin administered, the metabolism of the individual and other co-administrated medications<sup>5,6</sup>.
- I. Because the APTT can vary with technique, method, equipment, reagent lot and heparin used, each laboratory must establish its own therapeutic ranges, or verify them whenever one or more of the aforementioned variables is changed. This can be done by simultaneously determining the APTT and the heparin concentration for samples from patients receiving heparin therapy. A doseresponse curve can be calculated from the data using regression analysis, and the APTT range corresponding to a heparin concentration of 0.3 to 0.7 U/mL (by a factor Xa inhibition assay) can be derived<sup>4,5,6</sup>.

## Internal Quality Control

|                     |                                                                                                                        |
|---------------------|------------------------------------------------------------------------------------------------------------------------|
| Normal range:       | Dade® Ci-Trol® 1, Ci-Trol <b>CONTROL 1</b> , or<br><b>CONTROL N</b>                                                    |
| Pathological range: | Dade® Ci-Trol® 2, Ci-Trol <b>CONTROL 2</b> , or<br>Dade® Ci-Trol® 3, Ci-Trol <b>CONTROL 3</b> , or<br><b>CONTROL P</b> |
| Heparin monitoring: | Ci-Trol <b>HEPARIN CONTROL 1</b><br>Ci-Trol <b>HEPARIN CONTROL 2</b>                                                   |

Two controls (one in the normal range and one in the pathological/therapeutical range) must be measured at the start of the test run, after each change of reagent vial, and at least once during an 8-hour shift. The control material should be prepared and processed in the same manner as the patient samples. Each laboratory should establish its own confidence intervals for the controls. This interval is generally  $\pm 2$  to  $\pm 2.5$  standard deviations (SD) from the mean control value. If the control values are outside of the confidence interval, the controls, reagents and instrument must be checked. Before reporting the patient values, it is recommended that all steps should be documented that were taken to identify and rectify the problem. New control ranges should be defined for each new lot of reagents or controls.

## Results

Results of the activated partial thromboplastin time testing should be reported as the APTT in seconds. These results should be related to the normal range for APTT testing in each laboratory. It is suggested that the patient results be reported to the clinician in conjunction with the normal range. Control values for the reagent test system should never be used in place of a normal range. Furthermore, the reporting of APTT results in terms of an upper normal only may result in incorrect interpretation. Shortened APTT results may also indicate some abnormal condition in the patient's coagulation system.

## Limitations

APTT testing encompasses the entire clotting process from contact activation to fibrin formation and is therefore more susceptible to variations than specific individual tests. The control and use of APTT is therefore subject to inherent limitations. Control of plasma sample conditions is strictly emphasized. Studies have shown that sample decomposition may occur more rapidly in stored samples that are not refrigerated. Extremely small plasma volumes (prior to testing) are to be avoided since pH changes in the plasma from physiological conditions may be encountered. Such changes may lead to the decomposition of plasma components of the blood coagulation system. It should be noted that APTT time testing may be affected by a number of commonly administered drugs. Decrease in time of APTT determination in conjugated estrogen therapy in males and oral contraceptive administration in females has been reported<sup>13,14</sup>. Increase in the APTT has been seen in diphenylhydantoin, heparin, warfarin, naloxone and radiographic agent administration<sup>15,16</sup>. Therapeutic doses of hirudin or other direct thrombin inhibitors may prolong clotting times<sup>17</sup>.

Lipoglycopeptide antibacterial drugs (such as oritavancin or telavancin) may interfere with APTT based assays. Consult Instructions for Use of respective drugs.

In addition, the choice of anticoagulant (i.e. citrate vs. oxalate) and the condition of the specimen (e.g. hemolyzed, lipemic, parenteral feeding, etc.) may affect results<sup>6,18,19</sup>.

The latter is particularly true of optical instrumentation measurements of the APTT. Blood clotting factor deficiencies which should produce prolonged clotting times may be compensated for or made to appear normal by elevated levels of one or more different clotting factors. Similarly, the presence of active intermediates which would tend to reduce the clotting time may also mask conditions that would normally lead to prolongation of the APTT. Mild or minor deficiencies in several factors may have an additive effect on increasing the APTT.

The APTT assay is a functional test which screens for global coagulation disorders of the endogenous coagulation system. It is common knowledge that a low concentration of coagulation factors: FXII, FV, FX and high concentration of fibrin(ogen) degradation products also have influence on the assay<sup>19</sup>.

Unexpected abnormal APTT results should always be followed by additional coagulation studies to determine the source of abnormal results.

Action of heparin as an anticoagulant is related to its ability in conjunction with a plasma cofactor to interfere with several aspects of the coagulation mechanism, thus retarding the rate of fibrin formation (see "Monitoring of Unfractionated Heparin Therapy with APTT", page 4). Heparinase (Hepzyme) can be used as a heparin neutralizer in plasma to rule out heparin contamination in coagulation testing<sup>21</sup>.

The manufacturer has validated use of these reagents on various analyzers to optimize product performance and meet product specifications. Please note that the applications on other analyzers can be validated by the instrument manufacturer in accordance with the requirements of the REGULATION (EU) 2017/746 under their responsibility as long as the intended purpose and performance are not modified. User defined modifications are not supported by the manufacturer as they may affect performance of the system and assay results. It is the responsibility of the user to validate modifications to these instructions or use of the reagents on analyzers other than those included in Application Sheets or these Instructions for Use.

Results of this test should always be interpreted in conjunction with the patient's medical history, clinical presentation and other findings.

## Expected Values

In a study of ostensibly healthy individuals using a specific lot of **ACTIN**, the following values were obtained:

|                | n   | Median<br>[s] | 90 % Reference Interval           |                                    |
|----------------|-----|---------------|-----------------------------------|------------------------------------|
|                |     |               | 5 <sup>th</sup> Percentile<br>[s] | 95 <sup>th</sup> Percentile<br>[s] |
| CA-1500 System | 111 | 26.6          | 22.7                              | 31.8                               |
| BCS® System    | 111 | 26.1          | 22.8                              | 31.0                               |

Reference ranges for other populations such as pediatric groups should also be established where warranted.

Reference intervals vary from laboratory to laboratory depending on the population served and the technique, method, equipment and reagent lot used. Therefore, each laboratory must establish its own reference intervals or verify them whenever one or more of the aforementioned variables are changed.

**Note:** CLSI Document C28-A2 (cited in H47-A)<sup>22,23</sup> states that a parametric approach (mean  $\pm$  2 SD) can be applied. The assumption of this approach (Gaussian normal distribution) must however be checked.

## Performance Characteristics

### Measuring Range

The measuring range depends on the individual application of the assay due to instrument related conditions. Application specific performance data are listed in the respective Reference Guides of the instruments.

### Sensitivity

#### Factor Sensitivity of **ACTIN**

According to CLSI H47-A2, the APTT reagent/instrument combination used should provide abnormally prolonged results for plasmas that have less than 30 % of Norm factor activity of the coagulation factors VIII, IX and XI. CLSI H47-A2 recommends to determine sensitivity levels by serial dilution of normal plasma into deficient plasma. Sensitivity levels determined by this method should ideally be within 30 and 45 % of Norm. However, the factor sensitivity levels determined by this method is strongly dependent on deficient plasma used<sup>28</sup>.

#### Heparin Sensitivity

For determination of heparin sensitivity 64 samples from patients receiving unfractionated heparin were analyzed. By comparing the APTT obtained with **ACTIN** and the Heparin activity measured with HemosIL Liquid Anti Xa a correlation coefficient of  $r = 0.592$  was observed. Each individual laboratory hospital should determine its own therapeutic heparin range using the ex vivo method according the CLSI guidelines (H57-A2)<sup>29</sup>.

#### Lupus Sensitivity

By testing 25 samples with a confirmed lupus anticoagulant with **ACTIN**, the APTT of all samples did exceed the 99<sup>th</sup> percentile of the Reference Interval, demonstrating a medium to high sensitivity of **ACTIN** for the presence of lupus anticoagulants<sup>30</sup>.

### Precision

Precision studies using the methodologies listed in this insert show that properly performed APTT tests should result in a standard deviation (SD) which corresponds to a coefficient of variation (CV) of < 5 % in the normal range. Additional studies demonstrating reproducibility between duplicate determinations on the same sample indicate that the APTT should agree within 4 % in the normal range when performed properly.

Other system specific results are given in the respective Reference Guides (Application Sheets).

The reproducibility was assessed by the manufacturer for APTT with Dade® Actin® Activated Cephaloplastin Reagent based on publicly available proficiency testing information in 2018. The overall reproducibility median CV% was found to be <5 % including lot, instrument, laboratory and operator variability factors.

## Technical Assistance

For customer support, contact your local technical support provider or distributor.

### Current Version of Application Sheets

**ACTIN** can be used in combination with various automated coagulation analyzers. Sysmex provides Reference Guides/Application Sheets for the coagulation analyzers listed in section "Materials Required but not Provided", page 3 under the dedicated link below:

[sysmex-ifu.com/ag](http://sysmex-ifu.com/ag)

As the manufacturer continuously monitors the product performance and safety, the users are required to ensure that they work with the correct revision of the instructions for the product lots in use. Please periodically review the availability of new electronic labeling revisions to ensure safe use of the product.

The IFU version number is visible on each product box label. Sysmex ensures that all products lots bearing the same IFU version number are compatible with the electronic labeling provided via [sysmex-ifu.com](http://sysmex-ifu.com).

## References

1. Kamal AH, Tefferi A, Pruthi RK. How to interpret and pursue an abnormal prothrombin time, activated partial thromboplastin time, and bleeding time in adults. *Mayo Clin Proc.* 2007;82:864-73.
2. Leung LLK. Perioperative evaluation of bleeding diathesis. *Hematology Am Soc Hematol Educ Program.* 2006;457-61.
3. Eikelboom JW, Hirsh J. Monitoring unfractionated heparin with the aPTT: Time for a fresh look. *Thromb Haemost.* 2006;96:547-52.
4. Bates SM, Johnston M, Hirsh J, et al. Use of a fixed activated partial thromboplastin time ratio to establish a therapeutic range for unfractionated heparin. *Arch Intern Med.* 2001;161:385-91.
5. Hirsh J, Raschke R. Heparin and low-molecular-weight heparin. The seventh ACCP Conference on Antithrombotic and Thrombolytic Therapy. *Chest.* 2004;126:188S-203S.
6. Marlar RA, Clement B, Gausman J. Activated partial thromboplastin time monitoring of unfractionated heparin therapy: Issues and recommendations. *Semin Thromb Hemost.* 2017;43:253-60.
7. Teruya J, West AG, Suell MN. Lupus Anticoagulant Assays. Questions answered and to be answered. *Arch Pathol Lab Med.* 2007;131:885-9.
8. Fritsma GA, Dembitzer FR, Randhawa A, et al. Recommendations for appropriate activated partial thromboplastin time reagent selection and utilization. *Am J Clin Pathol.* 2012;137:904-8.
9. Kershaw G. Performance of Activated Partial Thromboplastin Time (APTT): Determining Reagent Sensitivity to Factor Deficiencies, Heparin, and Lupus Anticoagulants. *Methods Mol Biol.* 2017;1646:75-83.
10. Levy JH, Szlam F, Wolberg AS, et al. Clinical use of the activated partial thromboplastin time and prothrombin time for screening: a review of the literature and current guidelines for testing. *Clin Lab Med.* 2014;34:453-77.
11. CLSI. Collection, Transport, and Processing of Blood Specimens for Testing Plasma-Based Coagulation Assays and Molecular Hemostasis Assays. Approved Guideline – Fifth Edition. CLSI document **H21-A5** [ISBN 1-56238-657-3]. CLSI, 940 West Valley Road, Suite 1400, Wayne, PA 19087-1898 USA, 2008.
12. Hattersley PG, Hayse D. The effect of increased contact activation time on the activated partial thromboplastin time. *Am J Clin Pathol.* 1976;66:479-82.
13. Ambrus JL, Schimert G, Lajos TZ, et al. Effect of antifibrinolytic agents and estrogens on blood loss and blood coagulation factors during open heart surgery. *J Med.* 1971;2:65-81.
14. Prasad RN, Koh SC, Viegas OA, et al. Effects on hemostasis after two-year use of low dose combined oral contraceptives with gestodene or levonorgestrel. *Clin Appl Thromb Hemost.* 1999;5:60-70.
15. Solomon GE, Hilgartner MW, Kutt H. Coagulation defects caused by diphenylhydantoin. *Neurology.* 1972;22:1165-71.
16. Young DS, editor. *Effects of Drugs on Clinical Laboratory Tests*, 5th ed. Washington: AACC Press; 2000.
17. Greinacher A, Warkentin TE. The direct thrombin inhibitor hirudin. *Thromb Haemost.* 2008;99:819-29.
18. Soloway HB, Cox SP, Donahoo JV. Sensitivity of the activated partial thromboplastin time to heparin: effect of anticoagulant used for sample collection. *Am J Clin Pathol.* 1973;59:760-2.
19. Lippi G, Franchini M, Montagnana M, et al. Quality and reliability of routine coagulation testing: can we trust that sample? *Blood Coag Fibrinolysis.* 2006;17:513-9.
20. Thomas L, Chapter 16 in: *Clinical Laboratory Diagnostics (Labor und Diagnose)*. 8th ed. TH Books; 2016.
21. Harenberg J, Reichel T, Malsch R, et al. Multicentric evaluation of heparinase on aPTT, thrombin clotting time and a new PT reagent based on recombinant human tissue factor. *Blood Coagul Fibrinolysis.* 1996;7:453-458.
22. NCCLS. One-Stage Prothrombin Time (PT) Test and Activated Partial Thromboplastin Time (APTT) Test; Approved Guideline. NCCLS document **H47-A** [ISBN 1-56238-301-9]. NCCLS, 940 West Valley Road, Wayne, PA 19087, 1996

23. NCCLS. How to Define and Determine Reference Intervals in the Clinical Laboratory; Approved Guideline—Second Edition. NCCLS document **C28-A2** [ISBN 1-56238-406-6]. NCCLS, 940 West Valley Road, Suite 1400, Wayne, Pennsylvania 19087-1898, USA 2000
24. Adcock DM, Kressin DC, Marlar RA. Effect of 3.2 % vs 3.8 % sodium citrate concentration on routine coagulation testing. *Am J Clin Pathol.* 1997;107:105–110 doi: 10.1093/ajcp/ 107.1.105.
25. Marlar RA, Potts RM, Marlar AA. Effect on routine and special coagulation testing values of citrate anticoagulant adjustment in patients with high hematocrit values. *Am J Clin Pathol.* 2006;126:400-5. doi: 10.1309/RRQKT2JEYV33D19D.
26. Adcock DM, Kressin DC, Marlar RA. The effect of time and temperature on routine coagulation tests. *Blood Coagul Fibrinolysis.* 1998;9:463-70.
27. Woodhams B, Giradot O, Blanco M, et al. Stability of coagulation proteins in frozen plasma. *Blood Coagul Fibrinolysis.* 2001;12:229-36.
28. Lawrie AS, Kitchen S, Efthymiou M, et al. Determination of APTT factor sensitivity- the misleading guideline. *Int J Lab Hematol.* 2013;35:652-657.
29. Geens T, Vertessen F, Malfait R, et al. Validation of the Sysmex CS5100 coagulation analyzer and comparison to the Stago STA-R analyzer for routine coagulation parameters. *Int J Lab Hematol.* 2015;37:372-81.
30. Wissel T, Wagner C. Lupus anticoagulants sensitivity of Siemens APTT reagents. *J Thromb Haemost.* 2013;Suppl.2:1167.

## Definition of Symbols

The following symbols may appear on the product labeling:

|          |                                                     |  |                                                                                          |
|----------|-----------------------------------------------------|--|------------------------------------------------------------------------------------------|
|          | Do not reuse                                        |  | Use By                                                                                   |
|          | Batch Code                                          |  | Catalogue Number                                                                         |
|          | Caution                                             |  | Manufacturer                                                                             |
|          | Authorized representative in the European Community |  | Authorized representative in Switzerland                                                 |
|          | Contains sufficient for <n> tests                   |  | Biological Risks                                                                         |
|          | <i>In Vitro</i> Diagnostic Medical Device           |  | Temperature Limitation                                                                   |
|          | Consult instruction for Use                         |  | Non-sterile                                                                              |
|          | CE marking of conformity                            |  | CE marking of conformity with notified body ID number. Notified body ID number can vary. |
|          | Contents                                            |  | Reconstitution volume                                                                    |
|          | Level                                               |  | Keep away from sunlight and heat                                                         |
|          | Warning                                             |  | Danger                                                                                   |
|          | Prescription device (US only)                       |  | Device Identification (UDI) barcode                                                      |
|          | REACH Authorization Number                          |  |                                                                                          |
| xx/xx/xx |                                                     |  |                                                                                          |

## **Legal Information**

Actin, BCS, Ci-Trol and Dade are trademarks of Siemens Healthineers.

Sysmex is a trademark of SYSMEX CORPORATION.

All other trademarks are the property of their respective owners.

© Siemens Healthineers, 2008–2024. All rights reserved.



**Siemens Healthcare Diagnostics Products GmbH**

Emil-von-Behring-Str. 76

35041 Marburg

Germany

[siemens-healthineers.com](https://www.siemens-healthineers.com)